



PIPELINE INSPECTION CHEMICAL/BACTERIA KIT - SOLIDS

MIC Corrosion Areas, Mud, Soil, Solids, Pipeline Dust, Scale, Film, and Slime

ON-SITE ANALYSIS

KIT CONTENTS

CHEMICAL KIT

- 1 - Distilled Water Tube, Labeled "DI Water" - Blue Cap
- 2 - 1 ml Plastic Droppers
- 1 - Hydrochloric Acid (2 N HCl) Tube - Yellow Cap
- 2 - Carbonate Indicator Paper Strips - Carbonate Tube
- 2 - Calcium Indicator Paper Strips - Calcium Tube
- 2 - Chloride Indicator Paper Strips - Chloride Tube
- 2 - Sulfate Indicator Paper Strips - Sulfate Tube
- 2 - Iron Indicator Paper Strips - Iron Tube
- 2 - Manganese Indicator Paper Strips - Manganese Tube
- 1 - Manganese Reagent No. 1 Tube - White Cap
- 1 - Manganese Reagent No. 2 Tube - Red Cap
- 2 - Nitrate Indicator Paper Strips - Nitrate Tube
- 2 - Nitrite Indicator Paper Strips - Nitrite Tube
- 2 - pH Indicator Paper Strips (Sealed Plastic Bag)
- 2 - Lead-Acetate Paper Strips with Test Cap (Sealed Plastic Bag)
- 1 - Sterile Sampling Spoon (Sealed Plastic Bag)
- 1 - Pack of 2 Sterile Cotton Tipped Swabs (Sealed Paper Bag)
- 1 - 6 mil plastic sampling bag
- 1 - Pair of Nitrile Gloves

BACTERIA KIT

- 1 - Anaerobic Digestion Solution Tube - Labeled "ADS" - Orange Cap
- 1 - Sterile Sampling Spoon (sealed plastic bag)
- 5 - Bottles Postgate Media "B" 2% Salt Sulfate Reducer Media - Green Cap
- 5 - Bottles of Anaerobic Phenol Red 2% Salt (Acid Producers - Anaerobic Media) - Red Cap
- 4 - 3-mL Syringes
- 2 - Alcohol Swabs
- 1 - Pair of Nitrile Gloves

INSTRUCTION SHEETS

- 1 - Chemical Kit Sampling and Testing Instructions, consisting of four (4) pages.
- 1 - Bacteria (Microbiology Survey) Kit Sampling and Testing Instructions consisting of three (3) pages

REPORTING FORMS

- 1 - Chemical Testing Reporting Form, consisting of one (1) page.
- 1 - Microbiology Reporting Form, consisting of two (2) pages.

WARRANTY

Products are warranted by MSES Consultants, Inc. Corrosion Products Division, to perform as described in the technical literature furnished with each product, provided the products are used in accordance with the directions provided. This kit must be used before the expiration date shown on the kit. Adequate quality control must be done by the user of the products.

MSES Consultants, Inc. Corrosion Products Division disclaims any implied warranty of merchantability or fitness of its products for any purpose other than that described in its technical literature and in no event shall MSES Consultants, Inc. Corrosion Products Division be liable for any consequential damages out of the aforesaid warranty.

If you have any questions about any of the products and services provided by MSES Consultants, Inc. Corrosion Products Division, please call or write to the address below:

PRODUCT INFORMATION QUESTIONS

Corrosion Products Division
MSES Consultants, Inc.
P. O. Drawer 190
Clarksburg, West Virginia 26302-0190
Office Phone: 304.624.9700
Fax: 304.622.0981
24 Hour Phone: 304.842.3325
E-mail: mSES@msesinc.com

ORDERING INFORMATION

Corrosion Products Division
MSES Consultants, Inc.
Attn: Internal Corrosion Department
P. O. Drawer 190
Clarksburg, West Virginia 26302-0190
Office Phone: 304.624.9700
Fax: 304.622.0981
E-mail: mSES@msesinc.com
Web Store: www.msESproducts.com